

## 6.9. Worksheet

### Determination of the absorbed dose to water in a high-energy photon beam

User: Patrícia Nicolucci Date: 25/09/2020

#### 1. Radiation treatment unit and reference conditions for $D_{w,Q}$ determination

Accelerator: Mevatron XII Nominal Acc Potential: 10 MV  
 Nominal dose rate: 300 MU min<sup>-1</sup> Beam quality,  $Q$  ( $TPR_{20,10}$ ): \_\_\_\_\_  
 Reference phantom: water Set up:  SSD  SAD  
 Reference field size: 10x10 cm x cm Reference distance (cm): 100  
 Reference depth  $z_{ref}$ : 10 g cm<sup>-2</sup>

#### 2. Ionization chamber and electrometer

Ionization chamber model: NE 2571 farmer Serial no.: XB3216  
 Chamber wall material: grafite thickness: 0,065 g cm<sup>-2</sup>  
 Waterproof sleeve material: latex thickness: 0,082 g cm<sup>-2</sup>  
 Phantom window material: PMMA thickness: 1,2 g cm<sup>-2</sup>

Absorbed-dose-to-water calibration factor <sup>a</sup>  $N_{D,w,Q_0} =$   $9,61 \times 10^{-3}$   Gy nC<sup>-1</sup>  Gy rdg<sup>-1</sup>

Calibration quality  $Q_0$ :  <sup>60</sup>Co  photon beam Calibration depth: 5 g cm<sup>-2</sup>

If  $Q_0$  is photon beam, give  $TPR_{20,10}$ : 0,568

Reference conditions for calibration  $P_0$ : 101,2 kPa  $T_0$ : 20 °C Rel. humidity: 60 %

Polarizing potential  $V_j$ : 300 V Calibration polarity:  +ve  -ve  corrected for polarity effect

User polarity:  +ve  -ve

Calibration laboratory: IRD-CNEN Date: 23/03/2020

Electrometer model: PTW Unidos Serial no.: 880001

Calibrated separately from chamber:  yes  no Range setting: auto

If yes Calibration laboratory: \_\_\_\_\_ Date: \_\_\_\_\_

#### 3. Dosimeter reading <sup>b</sup> and correction for influence quantities

Uncorrected dosimeter reading at  $V_j$  and user polarity: 83,44  nC  rdg

Corresponding accelerator monitor units: 100 MU

Ratio of dosimeter reading and monitor units:  $M_I =$  \_\_\_\_\_  nC MU<sup>-1</sup>  rdg MU<sup>-1</sup>

(i) Pressure  $P$ : 101,2 kPa Temperature  $T$ : 22 °C Rel. humidity (if known): 48 %

$$k_{TP} = \frac{(273.2 + T) P_0}{(273.2 + T_0) P} = \frac{(273.2 + 22) \cdot 101.2}{(273.2 + 20) \cdot 101.2} = \frac{295.2}{293.2} = 1.0068$$

(ii) Electrometer calibration factor <sup>c</sup>  $k_{elec}$ :  nC rdg<sup>-1</sup>  dimensionless  $k_{elec} =$  \_\_\_\_\_

(iii) Polarity correction <sup>d</sup> rdg at  $+V_j$ :  $M_+$  = \_\_\_\_\_ rdg at  $-V_j$ :  $M_-$  = 83,47

$$k_{pol} = \frac{|M_+| + |M_-|}{2M} = \frac{83.44 + 83.47}{2 \cdot 83.44} = 1.0001$$

(iv) Recombination correction (two-voltage method)

Polarizing voltages:  $V_1$  (normal) = \_\_\_\_\_ V  $V_2$  (reduced) = 150 V

Readings<sup>a</sup> at each V:  $M_1$  = \_\_\_\_\_  $M_2$  = 83,33

Voltage ratio  $V_1 / V_2$  = \_\_\_\_\_ Ratio of readings  $M_1 / M_2$  = \_\_\_\_\_

Use Table 4.VII for a beam of type:  pulsed  pulsed-scanned

$a_0$  = \_\_\_\_\_  $a_1$  = \_\_\_\_\_  $a_2$  = \_\_\_\_\_

$$k_s = a_0 + a_1 \left( \frac{M_1}{M_2} \right) + a_2 \left( \frac{M_1}{M_2} \right)^2 = \text{_____} \quad \text{f, g}$$

Corrected dosimeter reading at the voltage  $V_1$ :

$$M_Q = M_1 k_{TP} k_{elec} k_{pol} k_s = \text{_____} \quad \square \text{ nC MU}^{-1} \quad \square \text{ rdg MU}^{-1}$$

**4. Absorbed dose to water at the reference depth,  $z_{ref}$**

Beam quality correction factor for user quality  $Q$ :  $k_{Q,Q_0}$  = \_\_\_\_\_

taken from  Table 6.III  Other, specify: \_\_\_\_\_

$$D_{w,Q}(z_{ref}) = M_Q N_{D,w,Q_0} k_{Q,Q_0} = \text{_____ Gy MU}^{-1}$$

**5. Absorbed dose to water at the depth of dose maximum,  $z_{max}$**

Depth of dose maximum:  $z_{max}$  = 2,5 g cm<sup>-2</sup>

(i) SSD set-up

Percentage depth-dose at  $z_{ref}$  for a 10 cm x 10 cm field size:  $PDD(z_{ref} = \text{10 g cm}^{-2}) = \text{80,1} \%$

Absorbed-dose calibration of monitor at  $z_{max}$ :

$$D_{w,Q}(z_{max}) = 100 D_{w,Q}(z_{ref}) / PDD(z_{ref}) = \text{_____ Gy MU}^{-1}$$

(ii) SAD set-up

TMR at  $z_{ref}$  for a 10 cm x 10 cm field size:  $TMR(z_{ref} = \text{_____ g cm}^{-2}) = \text{_____}$

Absorbed-dose calibration of monitor at  $z_{max}$ :

$$D_{w,Q}(z_{max}) = D_{w,Q}(z_{ref}) / TMR(z_{ref}) = \text{_____ Gy MU}^{-1}$$

<sup>a</sup> Note that if  $Q_0$  is <sup>60</sup>Co,  $N_{D,w,Q_0}$  is denoted by  $N_{D,w}$ .

<sup>b</sup> All readings should be checked for leakage and corrected if necessary

<sup>c</sup> If the electrometer is not calibrated separately set  $k_{elec} = 1$

<sup>d</sup>  $M$  in the denominator of  $k_{pol}$  denotes reading at the user polarity. Preferably, each reading in the equation should be the average of the ratios of  $M$  (or  $M_+$  or  $M_-$ ) to the reading of an external monitor,  $M_{em}$ .

It is assumed that the calibration laboratory has performed a polarity correction. Otherwise  $k_{pol}$  is determined according to

rdg at  $+V_1$  for quality  $Q_0$ :  $M_- = \text{_____}$  rdg at  $-V_1$  for quality  $Q_0$ :  $M_+ = \text{_____}$

$$k_{pol} = \frac{\left[ \frac{|M_+| + |M_-|}{|M|} \right]_{Q_0}}{\left[ \frac{|M_+| + |M_-|}{|M|} \right]_{Q_0}} = \text{_____}$$

<sup>e</sup> Strictly, readings should be corrected for polarity effect (average with both polarities). Preferably, each reading in the equation should be the average of the ratios of  $M_1$  or  $M_2$  to the reading of an external monitor,  $M_{em}$ .

<sup>f</sup> It is assumed that the calibration laboratory has performed a recombination correction. Otherwise the factor  $k'_s = k_s / k_{s,Q_0}$  should be used instead of  $k_s$ . When  $Q_0$  is <sup>60</sup>Co,  $k_{s,Q_0}$  (at the calibration laboratory) will normally be close to unity and the effect of not using this equation will be negligible in most cases.

<sup>g</sup> Check that  $k_s - 1 \approx \frac{M_1/M_2 - 1}{V_1/V_2 - 1}$

<sup>h</sup> Note that if  $Q_0$  is <sup>60</sup>Co,  $k_{Q,Q_0}$  is denoted by  $k_Q$ , as given in Table 6.III.